

AccuDiag™
AAS Rat Kidney/Stomach
Tissue
IFA Kit

Cat# 250608D

IVD  See external Label  2 -8°C  Σ= 10x 8wells

Test	AAS Rat Kidney/Stomach Tissue IFA
Method	Indirect Fluorescent Antibody Method
Principle	Qualitative & Semi Quantitative
Sample	10 µL
Total Time	~ 80 min.
Shelf Life	12-18 Months
Sensitivity	100 %
Specificity	100 %

INTENDED USE

The Diagnostic Automation, Inc. (AAS) Rat Kidney/Stomach Tissue Test System is designed for the qualitative and semi-quantitative detection of antinuclear, mitochondrial, smooth muscle and parietal cell antibodies by the indirect fluorescent antibody IFA technique. It aids in determining SLE and differentiating clinically similar connective tissue disorders, and is for In Vitro diagnostic use.

SUMMARY AND EXPLANATION

AAS Rat Kidney/Stomach Tissue Test System combines the DAI IFA Mitochondrial Antibody Test System and the DAI IFA Smooth Muscle Antibody Test Systems into a single system, which enables one to monitor four primary autoantibodies in a single test. This Test System will simultaneously detect antinuclear, mitochondrial, smooth muscle, and parietal cell antibodies. The DAI IFA AAS Rat Kidney/Stomach Tissue Test System is designed for use in conjunction with the DAI IFA ANA, MA, and SMA specific Test Systems.

The IFA technique was adapted to antinuclear antibody testing by several investigators following the basic methods originally described by Coons. This method has been used extensively for detecting the presence of ANA in the sera of patients with systemic lupus erythematosus (SLE), and other clinically similar connective tissue disorders. In addition, ANA may be associated with numerous drug-induced lupus syndromes which clinically mimic the spontaneous form of SLE. ANA are primarily composed of IgG; however, IgA and IgM ANA may also be detected. It is now recognized that many sources of nuclear material may be employed as a substrate for ANA testing. There are several different patterns of nuclear fluorescence. These various patterns and the basis for them are as follows:

Homogeneous - (Solid Diffuse) - Diffuse staining of the entire nucleus due to antibodies reactive with DNA-Nucleoprotein-Histone complexes.

Peripheral - (Rim, Shaggy) - Staining of the nuclear membrane due primarily to antibodies directed against DNA.

Speckled - Specks of staining dispersed throughout the nucleus due to antibodies against extractable nuclear antigens, RNP, or Sm.

Nucleolar - Staining of the nucleolar membranes due to antibodies reactive with RNA - nucleoprotein complexes. Although the level of ANA may not correlate with the clinical course of a particular autoimmune disease state, the various patterns of nuclear staining may be associated with specific disease states.

Mitochondrial - antibodies (MA) are found predominantly in patients (84% - 100%) with Primary Biliary Cirrhosis (PBC) and only occasionally (10% or less) in patients with chronic active hepatitis, cryptogenic cirrhosis and other diseases. For this reason, tests for MA have been recommended as a substitute for surgical exploration to provide confirmatory evidence when the diagnosis of PBC is suggested by clinical and laboratory features or histological findings. This recommendation is further supported by the absence of MA in extrahepatic biliary obstruction. Although the exact etiology of PBC has not been determined, the association of PBC with certain tissue antibodies, particularly MA, is well established. It has been suggested that PBC, chronic active hepatitis, and cryptogenic cirrhosis are various manifestations of a common autoimmune process. This was based on the observation that similar autoantibodies may be detected in the serum of all three diseases. Recent reports suggest that HBsAg may be the etiologic agent for PBC and possibly other liver diseases.

Smooth Muscle - antibodies (SMA) were first described by Johnson, et al and were thought to be specific for chronic active hepatitis. Although SMA are found in more than 50% of patients with chronic active hepatitis, they have also been found in association with PBC, asthma, and certain malignancies. SMA titers of 1:80 or greater that persist for several months to years are characteristically found in chronic active hepatitis. Patients with viral hepatitis on the other hand, rarely have titers above 1:40, and only have transient trace amounts of SMA. The specific antigen for SMA appears to be actin or actin-like substances which may be present in liver cells. Until this report, it was difficult to reconcile the presence of SMA with chronic active liver disease. Another report has shown SMA to be an autoantibody reactive with actin, the contractile substance of platelets, brush borders of epithelial cells, and other substances.

Parietal-Cell - antibodies (PCA) are seen in 90% of pernicious anemia patients. The test is helpful in differentiating this anemia from other macrocytic anemias. The parietal-cell antibody is seen in a large percentage of cases of atrophic gastritis and noted in a significant percentage of patients with iron deficiency anemia, thyroid disease, idiopathic Addison's disease, and juvenile diabetes mellitus. About one-third of the patients with thyroiditis or Graves' disease have parietal-cell antibodies. In normal subjects, parietal-cell antibodies are rare under the age of 20 years. There is an increasing incidence with age in females and males, which reflects increased frequency of atrophic gastritis. Intrinsic factor antibodies are usually of the IgG class and are found in 50 - 70% of patients with pernicious anemia. The intrinsic factor antibody is rarely seen in the absence of pernicious anemia.

TEST PRINCIPLE

The Diagnostic Automation, Inc. AAS Rat Kidney/Stomach Tissue Test System is a pre-standardized assay designed to screen patient sera for antinuclear, mitochondrial, smooth muscle and parietal-cell antibodies utilizing a single test procedure. The assay employs stomach and kidney tissue substrate sections in each well of an eight-well Slide. Antibodies are then diluted using goat anti-human immunoglobulin Conjugate adjusted for optimum use dilution with minimum background staining. The reaction occurs in two steps:



1. The first one involves the interaction of antibody in the patient's sera with the antigen on the Slide. In a positive specimen, antibodies in the serum will bind to the tissue section and remain attached after rinsing.
2. The second is the reaction between the Conjugate and the antigen-antibody reaction that produces an apple-green staining in a positive assay (see Assay Procedure).

The DAI AAS Rat Kidney/Stomach Tissue IFA Test System should be used to screen patients suspected of having SLE or other connective tissue diseases, autoimmune liver disease, such as chronic active hepatitis, or primary biliary cirrhosis, patients with pernicious anemia, and patients with symptoms consistent with possible autoimmune disease.

ANA: In a positive assay, the antinuclear antibody in the patient's sera interacts with the kidney and stomach nuclei. With the addition of the FITC Conjugate, an apple-green staining will occur. Antinuclear antibodies will exhibit a homogeneous, rim, speckled, or nucleolar pattern.

MA: In a positive assay, the mitochondrial antibody in the patients sera interacts with the mitochondrial antigens localized in the kidney proximal and more intensely, in distal tubular epithelium and gastric (stomach) parietal cells. With the addition of the FITC Conjugate, an apple-green staining will occur within the above structures.

SMA: In a positive assay, the smooth muscle antibody in the patient's sera interacts with the smooth muscle antigen in the muscularis band basal to the glandular mucosa of the stomach and in the smooth muscle tissue in the blood vessel walls. With the addition of the FITC Conjugate, a positive reaction is indicated by an apple-green staining within the muscularis band and blood vessel walls.

PCA: In a positive assay, the parietal-cell antibody in the patient's sera interacts with the gastric (stomach) parietal cells and not with the kidney proximal or distal tubular epithelium. With the addition of the FITC Conjugate, an apple-green staining will occur.

SPECIMEN COLLECTION AND PREPARATION

1. DAI recommends that the user carry out specimen collection in accordance with CLSI document M29: Protection of Laboratory Workers from Occupationally Acquired infectious Diseases. No known test method can offer complete assurance that human blood samples will not transmit infection. Therefore, all blood derivatives should be considered potentially infectious.
2. Only freshly drawn and properly refrigerated sera obtained by approved aseptic venipuncture procedures should be used in this assay. No anticoagulants or preservatives should be added. Avoid using hemolyzed, lipemic, or bacterially contaminated sera.
3. Store sample at room temperature for no longer than 8 hours. If testing is not performed within 8 hours, sera may be stored between 2-8°C for no longer than 48 hours. If delay in testing is anticipated, store test sera at -20°C or lower. Avoid multiple freeze/thaw cycles that may cause loss of antibody activity and give erroneous results.

MATERIALS AND COMPONENTS

Materials provided with the test kits

Each Test System contains the following components in sufficient quantities to perform the number of tests indicated on packaging label. **NOTE: Conjugate and Controls contain a combination of Proclin (0.05% v/v) and Sodium Azide (<0.1% w/v) as preservatives. Sorbent contains Sodium Azide (<0.1% w/v) as a preservative.**

Reactive Reagents

1. **Rat Kidney/Stomach Tissue Substrate Slides:** Ten, 8 -well Slides with absorbent blotter and desiccant pouch.
2. **Conjugate:** Goat anti-human immunoglobulin (polyvalent) labeled with fluorescein isothiocyanate (FITC). Contains phosphate buffer with BSA and counterstain. One, 3.5 mL, clear capped bottle. Ready to use.
3. **ANA Positive Control (Human Serum):** Will produce homegenous staining of the kidney substrate. One, 0.5mL, red-capped, vial. Ready to use.

4. **MA Positive Control (Human Serum):** Will produce mitochondrial staining of the kidney substrate. One, 0.5mL, blue- capped, vial. Ready to use.
5. **SMA Positive Control (Human Serum):** Will produce staining of the stomach smooth muscle substrate. One, 0.5mL, orange-capped, vial. Ready to use.
6. **PCA Positive Control (Human Serum):** Will produce staining of stomach substrate. One, 0.5mL, purple-capped, vial. Ready to use.
7. **Negative Control (Human Serum):** Will produce no detectable ANA, MA, SMA or PCA staining of the stomach or kidney substrate. One, 0.5mL, green-capped, vial. Ready to use.
8. **Diluent:** One, 30mL, green-capped, bottle containing phosphate-buffered-saline. Ready to use. Note: The Diluent will change color when combined with serum.
9. **Phosphate-buffered-saline (PBS):** pH 7.2 ± 0.2. Empty contents of each buffer packet into one liter of distilled or deionized water. Mix until all salts are thoroughly dissolved. Four packets, sufficient to prepare 4 liters.
10. **Mounting media (Buffered Glycerol):** Two, 3.0 mL, white-capped, dropper tipped vials.

Note: Kit also contains:

1. Component list containing lot specific information is inside the kit box.
2. Package insert providing instructions for use.

Materials required but not provided

1. Small serological, Pasteur, capillary, or automatic pipettes.
2. Disposable pipette tips.
3. Small test tubes, 13 x 100mm or comparable.
4. Test tube racks.
5. Staining dish. A large staining dish with a small magnetic mixing set-up provides an ideal mechanism for washing slides between incubation steps. Cover slips, 24x60mm, thickness No. 1.
7. Distilled or deionized water.
8. Properly equipped fluorescence microscope.
9. 1 Liter Graduated Cylinder.
10. Laboratory timer to monitor incubation steps.
11. Disposal basin and disinfectant (i.e: 10% household bleach – 0.5% Sodium Hypochlorite).

The following filter systems or their equivalent have been found to be satisfactory for routine use with transmitted or incident light darkfield assemblies:

TRANSMITTED LIGHT		
Light Source: Mercury vapor 200W or 50W		
Excitation Filter	Barrier Filter	Red Suppression Filter
KP490	K510 or K530	BG38
BG12	K510 or K530	BG38
FITC	K520	BG38
Light Source: Tungsten – Halogen 100W		
KP490	K510 or K530	BG38

INCIDENT LIGHT			
Light Source: Mercury Vapor 200, 100, 50 W			
Excitation Filter	Dichroic Mirror	Barrier Filter	Red Suppression Filter
KP500	TK510	K510 OR K530	BG38
FITC	TK510	K530	BG38
Light Source: Tungsten – Halogen 50 and 100 W			
KP500	TK510	K510 or K530	BG38
FITC	TK510	K530	BG38



ASSAY PROCEDURE

1. Remove slides from storage and allow them to warm to room temperature (20-25°C.) Tear open the protective envelope and remove slides. **DO NOT APPLY PRESSURE TO FLAT SIDES OF PROTECTIVE ENVELOPE.**
2. Identify each well with the appropriate patient sera and Controls. **NOTE: The Controls are intended to be used undiluted.** Prepare a 1:20 dilution (e.g.: 10µL of serum + 190 µl of diluent) of each patient serum. **The Diluent will undergo a color change confirming that the specimen has been combined with the Diluent.**
 - a. Users may titrate the Positive Control to endpoint to serve as a semi-quantitative (1+Minimally Reactive) Control. In such cases, the Control should be diluted two-fold in the Diluent. When evaluated by DAI, an endpoint dilution is established and printed on the Positive Control vial (\pm one dilution). It should be noted that due to variations within the laboratory (equipment, etc.), each laboratory should establish its own expected end-point titer for each lot of Positive Control.
 - b. When titrating patient specimens, initial dilutions should be prepared in Diluent and all subsequent dilutions should be prepared in the Diluent or PBS only.
3. With suitable dispenser (listed above), dispense 20µL of each Control and each diluted patient sera in the appropriate wells.
4. Incubate Slides at room temperature (20 - 25°C) 30 minutes.
5. Gently rinse Slides with PBS. **Do not direct a stream of PBS into the test wells.**
6. Wash Slides for two, 5 minute intervals, changing PBS between washes.
7. Remove Slides from PBS one at a time. Invert Slide and key wells to holes in blotters provided. Blot Slide by wiping the reverse side with an absorbent wipe. **CAUTION:** Position the blotter and Slide on a hard, flat surface. Blotting on paper towels may destroy the Slide matrix. **Do not allow the Slides to dry during the test procedure.**
8. Add 20µL of Conjugate to each well.
9. Repeat steps 4 through 7.
10. Apply 3 - 5 drops of Mounting Media to each Slide (between the wells) and coverslip. Examine Slides immediately with an appropriate fluorescence microscope.

NOTE: If delay in examining Slides is anticipated, seal coverslip with clear nail polish and store in refrigerator. It is recommended that Slides be examined on the same day as testing.

RESULTS

- **Negative:** Titers less than 1:20 are considered negative.
- **Positive:** A positive reaction is the presence of any pattern of nuclear apple-green staining observe at a 1:20 dilution based on a 1+ to 4+ scale of staining intensity. 1 + is considered a weak reaction and a 4 + a strong reaction. All sera positive at 1:20 should be tittered to endpoint dilution. This is accomplished by making a 1:20, 1:40, 1:180, etc. serial dilution of all positives. The endpoint titer is the highest dilution that procedures a 1+ positive reaction. (See Test Principle)
- Antinuclear, mitochondrial, smooth muscle, and parietal cell antibody reactions may be observed with this substrate.

PERFORMANCE CHARACTERISTICS

The Diagnostic Automation, Inc. AAS Rat Kidney/Stomach Tissue Test System was tested in parallel with a reference procedure as follows:

1. Routine ANA testing was performed by both procedures on 434 patient specimens. Of these 434 sera, 116 were positive by both procedures. The DAI AAS Rat Kidney/Stomach Tissue Test System showed 97% agreement with respect to positive and negative results, and 100% agreement with respect to staining patterns. Of the 29 discrepancies in titer, the DAI procedure was one

dilution lower in 16 specimens while the reference procedure was one dilution lower in 13 specimens. Of the 16 specimens with lower titers by the DAI procedure, all were one dilution discrepancies, and 13 of these 16 involved specimens that were negative by the DAI procedure and positive at 1:20 by the reference procedure.

2. Routine MA testing was performed by both procedures on 77 patient specimens. Of the 77 sera, 15 were positive by both procedures. The DAI IFA AAS Rat Kidney/Stomach/Liver Tissue Test System showed 100% agreement with respect to positive and negative results. Of the 15 positive MA sera, 13 were obtained from patients with a diagnosis of primary biliary cirrhosis and two low titer positives were obtained from patients who were undergoing routine employee health examinations.
3. Routine SMA testing was performed by both procedures on 69 serum specimens. Of these 69 sera, 28 were positive at a 1:40 or greater titer by both methods and 41 were negative. There were 6 discrepancies between the two methods with respect to titer. The DAI procedure was one dilution higher in four specimens and one dilution lower in two specimens. There were no discrepancies with respect to the number of negative sera.

QUALITY CONTROL

1. Every time the assay is run, a Positive Control a Negative Control and a Buffer Control must be included.
2. It is recommended that one read the Positive and Negative Controls before evaluating test results. This will assist in establishing the references required to interpret the test sample. If Controls do not appear as described below, results are invalid.
 - a. **Antinuclear Antibody:** Homogeneous Positive Control is characterized by diffuse staining of the entire nucleus in the kidney or stomach sections. The Negative Control is characterized by the absence of specific fluorescence and a red, or dull green, background staining of all the cells due to Evans Blue counterstain.
 - b. **Mitochondrial Antibody:** The Positive Control is characterized by apple-green staining in the proximal and distal tubular epithelium, and gastric parietal cells, with a staining intensity of 2+ to 4+. The Negative Control is characterized by the absence of fluorescent staining of the kidney cells.
 - c. **Smoothmuscle Antibody:** The Positive Control is characterized by apple-green fluorescent staining on the muscularis band of the stomach substrate. The Negative Control is characterized by the absence of fluorescent staining on the muscularis of the stomach muscle.
 - d. **Parietal Cell Antibody:** The Positive Control is characterized by a granular apple-green staining of the stomach parietal cells in columns that surround the smooth stomach muscularis band. The Negative Control is characterized by the absence of fluorescent staining in any of the stomach and kidney cells.
3. Additional Controls may be tested according to guidelines or requirements of local, state, and/or federal regulations or accrediting organizations.

NOTES:

- a. The intensity of the observed fluorescence may vary with the microscope and filter system used.
- b. Non-specific reagent trapping may exist. It is important to adequately wash slides to eliminate false positive results.
- c. Non-nuclear staining of the kidney and stomach substrate may be observed with some human sera. Report nuclear staining results only and disregard non-nuclear staining.

LIMITATIONS OF PROCEDURE

The DAI IFA AAS Rat Kidney/Stomach Tissue Test System is a laboratory diagnostic aid and by itself is not diagnostic. Positive test results may be found in diseases other than those described in the "Significance and Background" section of



this Package Insert. It is therefore imperative that positive test results be interpreted by a medical authority.

EXPECTED VALUES

The expected value in the normal population is negative or less than 1:20. However, apparently healthy individuals in the 5th to 7th decade of life may have positive results.

PRECAUTIONS


- For In Vitro Diagnostic Use.
- Follow normal precautions exercised in handling laboratory reagents. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing, gloves, and eye/face protection. Do not breathe vapor. Dispose of waste observing all local, state, and federal laws.
- The wells of the Slide do not contain viable organisms. However, consider the Slide **potentially bio-hazardous materials** and handle accordingly.
- The Controls are **potentially bio-hazardous materials**. Source materials from which these products were derived were found negative for HIV-1 antigen, HBsAg and for antibodies against HCV and HIV by approved test methods. However, since no test method can offer complete assurance that infectious agents are absent, these products should be handled at the Bio-safety Level 2 as recommended for any potentially infectious human serum or blood specimen in the Centers for Disease Control/National Institutes of Health manual "Biosafety in Microbiological and Biomedical Laboratories": current edition; and OSHA's Standard for Bloodborne Pathogens .
- Adherence to the specified time and temperature of incubations is essential for accurate results. **All reagents must be allowed to reach room temperature (20 - 25°C) before starting the assay.** Return unused reagents to their original containers immediately and follow storage requirements.
- Improper washing could cause false positive or false negative results. Be sure to minimize the amount of any residual PBS, by blotting, before adding Conjugate. Do not allow the wells to dry out between incubations.
- The Diluent, Conjugate, and Controls contain Sodium Azide at a concentration of <0.1% (w/v). Sodium Azide has been reported to form lead or copper azides in laboratory plumbing which may cause explosions on hammering. To prevent, rinse sink thoroughly with water after disposing of solution containing Sodium Azide. This preservative may be toxic if ingested.
- Dilution or adulteration of these reagents may generate erroneous results.
- Never pipette by mouth. Avoid contact of reagents and patient specimens with skin and mucous membranes.
- Avoid microbial contamination of reagents. Incorrect results may occur.
- Cross contamination of reagents and/or samples could cause erroneous results.
- Reusable glassware must be washed and thoroughly rinsed free of all detergents.
- Avoid splashing or generation of aerosols.
- Do not expose reagents to strong light during storage or incubation.
- Allowing the slide packet to equilibrate to room temperature prior to opening the protective envelope will protect the wells and blotter from condensation.
- Collect the wash solution in a disposal basin. Treat the waste solution with disinfectant (i.e.:10% household bleach - 0.5% Sodium Hypochlorite). Avoid exposure of reagents to bleach fumes.
- Do not expose any of the reactive reagents to bleach-containing solutions or to any strong odors from bleach-containing . Trace amounts of bleach (Sodium Hypochlorite) may destroy the biological activity of many of the reactive reagents within this Test System.
- Do not apply pressure to slide envelope. This may damage the substrate.
- The components of this Test System are matched for optimum sensitivity and reproducibility. Reagents from other manufacturers should not be interchanged. Follow Package Insert carefully.


- All components are stable until the expiration date printed on the label, provided the recommended storage conditions are strictly followed. Do not use beyond the expiration date. Do not freeze.
- Evans Blue Counterstain is a potential carcinogen. If skin contact occurs, flush with water. Dispose of according to local regulations.
- Do not allow slides to dry during the procedure. Depending upon lab conditions, it may be necessary to place slides in a moist chamber during incubations.

Storage Conditions

2-8°C	Unopened Test System
2-8°C	Mounting Media, Conjugate, Diluent, Slides, Positive and Negative Controls,
2-8°C	Rehydrated PBS (Stable for 30 days)
2-25°C	Phosphate-buffered-saline (PBS) Packets

**ISO 13485
ISO 9001**





Diagnostic Automation/Cortez Diagnostics, Inc.
23961 Craftsman Road, Suite E/F,
Calabasas, California 91302 USA

Date Adopted	Reference No. Cat # 250608D
2012-09-27	IFA – AAS Rat Kidney Stomach Tissue - 2013
EC REP	CEpartner4U, Esdoornlaan 13, 3951DB Maarn. The Netherlands. www.cepartner4u.eu
Revision Date: 02-06-2013	