


AccuDiag™
ASA
(Antiskin Antibody)
IFA Kit

Cat# 280806D



Test	ASA IFA
Method	Indirect Fluorescent Antibody Method
Principle	Qualitative & Semi Quantitative
Sample	10 µL
Total Time	~ 80 min.
Shelf Life	12-18 Months from the manufacturing date

INTENDED USE

The Diagnostic Automation, Inc. ASA (Antiskin Antibody) IFA Test System is designed for the qualitative and semi-quantitative detection of antibodies associated with Pemphigus and Bullous Pemphigoid by the indirect fluorescent antibody (IFA) technique, and is for In Vitro diagnostic use.

SUMMARY AND EXPLANATION

Most authorities agree that serum antibody studies of all cases of suspected Pemphigus Vulgaris (PV) and Bullous Pemphigoid (BP) are advisable. Specific antibodies are demonstrable in virtually all active cases, particularly if repeated tests are performed during different stages of the disease. The most frequently recommended substrate for these serum antibody studies is monkey esophagus. Both PV and BP specific antibodies can be detected in a single assay using this substrate. PV antibodies react with the intercellular substances between the squamous epithelium of monkey esophagus, and BP antibodies react with the basement membrane of the squamous epithelial cell layer (1, 4, and 6). Antinuclear antibody staining of the squamous epithelial cell nuclei may occur if serum is obtained from patients with systemic lupus or other connective tissue disorders. Distinct intercellular staining may be detected in cases of pemphigus vegetans, pemphigus foliaceus (including the Brazilian form), and pemphigus erythematosus. Although a single serological test is not sufficient for determining the dosage of drug therapy in PV, a two-fold drop in titer is an indication of effective control of the disease process. Although some correlation in titer and disease activity is found in BP, the prognostic value is less than in PV due to fluctuations in titer. Serum studies may be diagnostic of PV or BP; however, it is frequently advisable to perform direct immunohistochemical studies on biopsies of skin lesions. Direct immunofluorescent studies of skin biopsies may be facilitated with DAI Tissue Fixative and Wash Solution, prepared according to Michel. This new fixative eliminates the necessity of immediate freezing of the biopsy specimen and allows transport of fresh tissue biopsies of skin and other tissues up to 5 days at ambient temperatures.

TEST PRINCIPLE

The Diagnostic Automation, Inc. ASA (Antiskin Antibody) IFA Test System is designed to detect the presence of circulating pemphigus and bullous pemphigoid antibodies in human sera. The assay employs monkey esophagus tissue substrate and goat anti-human immunoglobulin adjusted for optimum use dilution and free of nonspecific background staining. The reaction occurs in two steps:

1. The first one involves the interaction of anti-skin antibodies in patient's sera with the monkey esophagus.
2. The second is the interaction of FITC-labeled anti-human immunoglobulin with antibodies attached to the intercellular cement of the epithelial cells in pemphigus, or the basement membrane zone in bullous pemphigoid in a positive assay (see Assay Procedure section for details).

The DAI ASA (Antiskin Antibody) IFA Test System will detect both PV and BP antibodies. The assay is a particularly useful laboratory aid in the diagnosis of PV and BP since the majority of untreated patients with active disease will contain anti-skin antibodies in their serum.

SPECIMEN COLLECTION AND PREPARATION

1. DAI recommends that the user carry out specimen collection in accordance with CLSI document M29: Protection of Laboratory Workers from Occupationally Acquired infectious Diseases. No known test method can offer complete assurance that human blood samples will not transmit infection. Therefore, all blood derivatives should be considered potentially infectious.
2. Only freshly drawn and properly refrigerated sera obtained by approved aseptic venipuncture procedures should be used in this assay. No anticoagulants or preservatives should be added. Avoid using hemolyzed, lipemic, or bacterially contaminated sera.
3. Store sample at room temperature for no longer than 8 hours. If testing is not performed within 8 hours, sera may be stored between 2-8°C for no longer than 48 hours. If delay in testing is anticipated, store test sera at -20°C or lower. Avoid multiple freeze/thaw cycles that may cause loss of antibody activity and give erroneous results. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine stability criteria for its laboratory (11).

MATERIALS AND COMPONENTS

Materials provided with the test kits

Each Test System contains the following components in sufficient quantities to perform the number of tests indicated on packaging label. **NOTE: Conjugate and Controls contain a combination of Proclin (0.05% v/v) and Sodium Azide (<0.1% w/v) as preservatives. Sorbent contains Sample Diluent Azide (<0.1% w/v) as a preservative.**

Reactive Reagents

1. **Monkey Esophagus Substrate Slides:** Ten, 6 - well Slides with absorbent blotter and desiccant pouch.
2. **Conjugate:** Goat anti-human immunoglobulin labeled with fluorescein isothiocyanate (FITC). Contains phosphate buffer with BSA and counterstain. One, 3.5 mL, amber capped bottle. Ready to use.
3. **PV Positive Control (Human Serum):** Will produce staining of the squamous epithelial cells of the substrate. One, 0.5mL, red-capped, vial. Ready to use.
4. **BP Positive Control (Human Serum):** Will produce staining of the basement membrane of the substrate. One, 0.5mL, blue-capped, vial. Ready to use.
5. **Negative Control (Human Serum)** Will produce no substrate staining. One, 0.5mL, green-capped, vial. Ready to use.



- Sample Diluent:** One, 30mL, green-capped, bottle containing phosphate-buffered-saline. Ready to use. Note: The Sample Diluent will change color when combined with serum.
- Phosphate-buffered-saline (PBS):** pH 7.2 ± 0.2. Empty contents of each buffer packet into one liter of distilled or deionized water. Mix until all salts are thoroughly dissolved. Four packets, sufficient to prepare 4 liters.
- Mounting media (Buffered Glycerol):** Two, 3.0 mL, white-capped, dropper tipped vials.

Note: Kit also contains:

- Component list containing lot specific information is inside the kit box.
- Package insert providing instructions for use.

Materials required but not provided

- Small serological, Pasteur, capillary, or automatic pipettes.
- Disposable pipette tips.
- Small test tubes, 13 x 100mm or comparable.
- Test tube racks.
- Staining dish. A large staining dish with a small magnetic mixing set-up provides an ideal mechanism for washing slides between incubation steps.
- Cover slips, 24x60mm, thickness No. 1.
- Distilled or deionized water.
- Properly equipped fluorescence microscope.
- 1 Liter Graduated Cylinder.
- Laboratory timer to monitor incubation steps.
- Disposal basin and disinfectant (i.e: 10% household bleach – 0.5% Sodium Hypochlorite).

The following filter systems or their equivalent have been found to be satisfactory for routine use with transmitted or incident light darkfield assemblies:

TRANSMITTED LIGHT		
Light Source: Mercury vapor 200W or 50W		
Excitation Filter	Barrier Filter	Red Suppression Filter
KP490	K510 or K530	BG38
BG12	K510 or K530	BG38
FITC	K520	BG38
Light Source: Tungsten – Halogen 100W		
KP490	K510 or K530	BG38

INCIDENT LIGHT			
Light Source: Mercury Vapor 200, 100, 50 W			
Excitation Filter	Dichroic Mirror	Barrier Filter	Red Suppression Filter
KP500	TK510	K510 OR K530	BG38
FITC	TK510	K530	BG38
Light Source: Tungsten – Halogen 50 and 100 W			
KP500	TK510	K510 or K530	BG38
FITC	TK510	K530	BG38

ASSAY PROCEDURE

- Remove slides from storage and allow them to warm to room temperature (20-25°C.) Tear open the protective envelope and remove slides. **DO NOT APPLY PRESSURE TO FLAT SIDES OF PROTECTIVE ENVELOPE.**
- Identify each well with the appropriate patient sera and Controls. **NOTE: The Controls are intended to be used undiluted.** Prepare a 1:10 dilution (e.g.: 10µL of serum + 90 µl of Sample Diluent or PBS) of each patient serum. **The Sample Diluent will undergo a color change confirming that the specimen has been combined with the Diluent.**

Dilution Options:

- Users may titrate the Positive Control to endpoint to serve as a semi-quantitative (1+ Minimally Reactive) Control. In such cases, the Control should be diluted two-fold in the Diluent. When evaluated by DAI, an endpoint dilution is established and printed on the Positive Control vial (± one dilution). It should be noted that due to variations within the laboratory (equipment, etc.), each laboratory should establish its own expected end-point titer for each lot of Positive Control.
 - When titrating patient specimens, initial dilutions should be prepared in Diluent and all subsequent dilutions should be prepared in the Diluent or PBS only.
- Wash Slides for 3-5 minutes in PBS.
 - Remove Slides from PBS and blot dry with six-well blotting paper. It is suggested that blotting paper be placed on a flat surface. Then place substrate Slide in an inverted position over the blotter. Press firmly on back of Slide. Do not allow tissue substrate to dry throughout the test procedure.
 - With suitable dispenser (listed above), dispense 20µL of each Control and each diluted patient sera in the appropriate wells.
 - Incubate Slides at room temperature (20 - 25°C) 30 minutes.
 - Gently rinse Slides with PBS. **Do not direct a stream of PBS into the test wells.**
 - Wash Slides for two, 5 minute intervals, changing PBS between washes.
 - Remove Slides from PBS one at a time. Invert Slide and key wells to holes in blotters provided. Blot Slide by wiping the reverse side with an absorbent wipe. **CAUTION:** Position the blotter and Slide on a hard, flat surface. Blotting on paper towels may destroy the Slide matrix. **Do not allow the Slides to dry during the test procedure.**
 - Add 20µL of Conjugate to each well.
 - Repeat steps 6 through 9.
 - Apply 3 - 5 drops of Mounting Media to each Slide (between the wells) and coverslip. Examine Slides immediately with an appropriate fluorescence microscope.

NOTE: If delay in examining Slides is anticipated, seal coverslip with clear nail polish and store in refrigerator. It is recommended that Slides be examined on the same day as testing.

RESULTS

- Any apple-green staining of the specific structures noted above on a scale of 1+ to 4+ is considered positive. A 1+ is considered a weak reaction, and 4+ a strong reaction. All sera positive at a 1:10 dilution should be titered to endpoint dilution. This is accomplished by making 1:20, 1:40, 1:80, etc., serial dilutions of all positives. The endpoint is the highest dilution that produces a discernible positive reaction. Specific nuclear staining of the epithelial cell nuclei is considered a positive test for antinuclear antibodies which may be associated with SLE and other connective tissue diseases.

- **Negative:** Titers less than 1:10 are considered negative.
- **Positive:**
 - Specific intercellular staining between the squamous epithelial cells is considered a positive test for pemphigus antibodies.
 - Specific basement membrane zone staining is considered a positive test for bullous pemphigoid.

QUALITY CONTROL

- Every time the assay is run, a Positive Control a Negative Control and a Buffer Control must be included.
- It is recommended that one read the Positive and Negative Controls before evaluating test results. This will assist in establishing the references required to interpret the test sample. If Controls do not appear as described below, results are invalid.



- a. **Negative Control** - characterized by the absence of fluorescent staining of the squamous epithelial cells, and/or basement membrane zone.
 - b. **Positive Control** – PV Positive Control is characterized by any apple-green fluorescent staining between the squamous epithelial cells. The BP Positive Control is characterized by specific staining of the basement membrane zone.
3. Additional Controls may be tested according to guidelines or requirements of local, state, and/or federal regulations or accrediting organizations.

NOTES:

- a. The intensity of the observed fluorescence may vary with the microscope and filter system used.
- b. Non-specific reagent trapping may exist. It is important to adequately wash slides to eliminate false positive results.

LIMITATIONS OF PROCEDURE

- 1. The DAI ASA (Antiskin Antibody) IFA Test System is a laboratory aid and by itself is not diagnostic.
- 2. The results should be interpreted in light of the patient’s clinical condition.
- 3. No definitive association between the pattern of fluorescence and any specific disease state is intended with this product.
- 4. No U.S. standard of potency.

PRECAUTIONS

- 1. For In Vitro Diagnostic Use.
- 2. Follow normal precautions exercised in handling laboratory reagents. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing, gloves, and eye/face protection. Do not breathe vapor. Dispose of waste observing all local, state, and federal laws.
- 3. The wells of the Slide do not contain viable organisms. However, consider the Slide **potentially bio-hazardous materials** and handle accordingly.
- 4. The Controls are **potentially bio-hazardous materials**. Source materials from which these products were derived were found negative for HIV-1 antigen, HBsAg and for antibodies against HCV and HIV by approved test methods. However, since no test method can offer complete assurance that infectious agents are absent, these products should be handled at the Bio-safety Level 2 as recommended for any potentially infectious human serum or blood specimen in the Centers for Disease Control/National Institutes of Health manual “Biosafety in Microbiological and Biomedical Laboratories”: current edition; and OSHA’s Standard for Bloodborne Pathogens.
- 5. Adherence to the specified time and temperature of incubations is essential for accurate results. **All reagents must be allowed to reach room temperature (20 - 25°C) before starting the assay.** Return unused reagents to their original containers immediately and follow storage requirements.
- 6. Improper washing could cause false positive or false negative results. Be sure to minimize the amount of any residual PBS, by blotting, before adding Conjugate. Do not allow the wells to dry out between incubations.
- 7. The Sample Diluent, Conjugate, and Controls contain Sodium Azide at a concentration of <0.1% (w/v). Sodium Azide has been reported to form lead or copper azides in laboratory plumbing which may cause explosions on hammering. To prevent, rinse sink thoroughly with water after disposing of solution containing Sodium Azide. This preservative may be toxic if ingested.
- 8. Dilution or adulteration of these reagents may generate erroneous results.
- 9. Never pipette by mouth. Avoid contact of reagents and patient specimens with skin and mucous membranes.
- 10. Avoid microbial contamination of reagents. Incorrect results may occur.
- 11. Cross contamination of reagents and/or samples could cause erroneous results.

- 12. Reusable glassware must be washed and thoroughly rinsed free of all detergents.
- 13. Avoid splashing or generation of aerosols.
- 14. Do not expose reagents to strong light during storage or incubation.
- 15. Allowing the slide packet to equilibrate to room temperature prior to opening the protective envelope will protect the wells and blotter from condensation.
- 16. Collect the wash solution in a disposal basin. Treat the waste solution with disinfectant (i.e.:10% household bleach - 0.5% Sodium Hypochlorite). Avoid exposure of reagents to bleach fumes.
- 17. Do not expose any of the reactive reagents to bleach-containing solutions or to any strong odors from bleach-containing. Trace amounts of bleach (Sodium Hypochlorite) may destroy the biological activity of many of the reactive reagents within this Test System.
- 18. Do not apply pressure to slide envelope. This may damage the substrate.
- 19. The components of this Test System are matched for optimum sensitivity and reproducibility. Reagents from other manufacturers should not be interchanged. Follow Package Insert carefully.
- 20. All components are stable until the expiration date printed on the label, provided the recommended storage conditions are strictly followed. Do not use beyond the expiration date. Do not freeze.
- 21. Evans Blue Counterstain is a potential carcinogen. If skin contact occurs, flush with water. Dispose of according to local regulations.
- 22. Do not allow slides to dry during the procedure. Depending upon lab conditions, it may be necessary to place slides in a moist chamber during incubations.



Storage Conditions

2-8°C	Unopened Test System
2-8°C	Mounting Media, Conjugate, Sample Diluent, Slides, Positive and Negative Controls,
2-8°C	Rehydrated PBS (Stable for 30 days)
2-25°C	Phosphate-buffered-saline (PBS) Packets

REFERENCES

- 1. Cooperative study. Uses for Immunofluorescence test of skin and sera. Utilization immunofluorescence in the diagnosis of Bullous diseases, Lupus Erythematosus, and certain other dermatoses. Arch. Dermatol. 111:372-381, 1975.
- 2. Jablonski S, Chorzelski TP, Beutner EH, et al: Indications for skin and serum immunofluorescent studies in dermatology. In: Beutner EH, Chorzelski TP, Bean S, et al (Eds): Immunopathology of the Skin. Stroudsburg, PA, Dowden, Hutchinson, and Ross, pp. 1-24, 1973.
- 3. Chorzelski TP, Jablonski S, Beutner EH: Clinical significance of pemphigus antibodies in: Beutner EH, Chorzelski TP, Bean S, et al (Eds): Immunopathology of the Skin, Stroudsburg, PA, Dowden, Hutchinson, and Ross. pp. 25-43, 1973.
- 4. Lever WF: Pemphigus and Pemphigoid. Springfield, IL, Charles C. Thomas, publisher. pp. 3-226, 1965.
- 5. Michel B, Milner Y, David K: Preservation of Tissue-Fixed immunoglobulins in skin biopsies of patients with lupus erythematosus and bullous diseases: Preliminary report. J. Invest. Dermatology 59: 449-454, 1973.
- 6. Anderson P, Hale WL: Immunohistology of human antibodies on monkey esophagus, In: Beutner EH, Chorzelski TP, Bean S, et al (Eds): Immunopathology of the Skin. Stroudsburg, PA, Dowden, Hutchinson, and Ross: pp. 271-286, 1973.
- 7. Tan EM, Vaughn JH: antinuclear antibodies: Significance of biochemical specificities, In Beutner EH, Chorzelski TP, Bean S, et al (Eds):Immunopathology of the Skin. Stroudsburg, PA, Dowden, Hutchinson, and Ross: pp. 367-378, 1973.

8. Procedures for the collection of diagnostic blood specimens by venipuncture - Second Edition: approved Standard (1984). Published by National Committee for Clinical Laboratory Standards.
9. Lennette DA: Collection and preparation of specimens for virological examination. In: Manual of Clinical Microbiology, 4th ed., EH Lennette, A Balows, WJ Hausler, and HJ Shadomy (Eds): American Society for Microbiology, Washington, DC. Ch. 61, pp. 687-693, 1985.
10. U.S. Department of Labor, Occupational Safety and Health Administration: Occupational Exposure to Bloodborne Pathogens, Final Rule. Fed. Register 56:64175-64182, 1991.
11. Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guidelines – 4th Edition (2010). CLSI Document GP44-A4 (ISBN 1-56238-724-3). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, PA 19087.

<p>ISO 13485 ISO 9001</p> 			
 <p>Diagnostic Automation/Cortez Diagnostics, Inc. 23961 Craftsman Road, Suite E/F, Calabasas, California 91302 USA</p>			
Date Adopted	Cat # 280806D		
2014-10-06	IFA – ASA (Antiskin Antibody) - 2015		
<table border="1"> <tr> <td style="padding: 2px;">EC</td> <td style="padding: 2px;">REP</td> </tr> </table>	EC	REP	<p>CEpartner4U, Esdoornlaan 13, 3951DB Maarn. The Netherlands. www.cepartner4u.eu</p>
EC	REP		
Revision B Date: 05-06-2015			