

Diagnostic Automation, Inc.


Test Background

The determination of serum or plasma levels of thyroid stimulating hormone (TSH or thyrotropin) is recognized as a sensitive method in the diagnosis of primary and secondary hypothyroidism.

Specifications

Principle	Sandwich Complex
Total Time	~80 min
Shelf Life	12 months from manufacturing date
Sample	50 µL

**Minimum
detectable
range:**
0.2 µIU/mL



AccuDiag™
TSH
ELISA

Intended for the quantitative determination of thyroid stimulating hormone (TSH) concentration in human serum.



Advantages:

- Specific and sensitive technology
- High Throughput Screening (HTS) compatible
- CE Marked
- Cost Efficient