



AccuDiag™ Milk ELISA Kit

REF 5165

PIC FD5165ZR44

IVD See External Label 2°C 8°C 96 Tests

Milk ELISA	
Principle	Indirect ELISA
Detection	Quantitative
Sample	100 µL Sample
Incubation Time	60 minutes
Sensitivity	0.05 ppm
Shelf Life	12 Months from the manufacturing date

PRODUCT FEATURES

- Very easy to use with little training
- Highly specific and consistent Assay
- Provides accurate results quickly
- Reading of results both visually and as absorbance data

INTENDED USE

The DAI Milk ELISA represents a highly sensitive detection system for milk proteins based on NIST 1549 reference material. The test is likewise capable of the quantification of casein and β-lactoglobulin residues in food and is validated for cookies, bread crumbs, sausage, orange juice, wine, soy products and chocolate.

SIGNIFICANCE AND SUMMARY

Bovine milk belongs to the most important allergenic food ingredients especially for children. Already very low amounts of bovine milk can cause

allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, milk allergic persons must strictly avoid the consumption of milk or milk containing food. In particular the presence of hidden milk proteins such as in sausage, cookies, convenience food or beverages represent a critical problem for milk allergic persons. According to EU Directive 2003/89/EG the addition of bovine milk has to be labeled. For the detection of bovine milk in foodstuffs, sensitive detection systems are required.

Approximately 80% of bovine milk proteins are caseins. β-Lactoglobulin, the major allergen of whey, represents further 10% of the total protein

ASSAY PRINCIPLE

The DAI Milk ELISA test is based on the principle of the enzyme linked immunosorbent as-say. An antibody mixture is bound on the surface of a microtiter plate. Milk protein containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody mixture directed against milk proteins is given into the wells and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue color. The color development is inhibited by the addition of a stop solution, and the color turns yellow. The yellow color is measured photo-metrically at 450 nm. The concentration of milk proteins is directly proportional to the color intensity of the test sample.

SPECIMEN COLLECTION & PREPARATION

Due to a high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be cleaned thoroughly before and after each sample. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for **solid** samples:

1. To maximize homogeneity and representative-ness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 0.5 g of the homogenized mixture is suspended in 10 mL of pre-diluted extraction and sample dilution buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. Due to high matrix effects meat and sausage samples should be further diluted 1 + 4 with pre-diluted extraction and sample dilution buffer.
5. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the pre-diluted extraction and sample dilution buffer is necessary. The additional dilution has to be con-sidered when calculating the concentration.

The following sample preparation should be applied for **liquid** samples: 0.5 mL of liquid sample is diluted in 9.5 mL of pre-diluted extraction and sample dilution buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes. The process is continued at point 3 of solid sample extraction process.



REAGENTS

Materials provided with the kit

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with milk protein binding antibodies.
2. Milk protein Standards, based on NIST RM 1549 reference material: 5 vials with 4.0 mL (0, 0.4, 1, 4, 10 ppm of milk protein), as 100x concentrate, dyed blue. Dilute 20 µL of standard with 1980 µL pre-diluted extraction and sample dilution buffer to achieve the concentrations named above. Stored at 4°C the diluted standards are stable for at least 24 hours.

Note: The concentrations above refer to the 100x diluted standards

3. Conjugate (anti-milk protein-peroxidase): 15 mL, dyed red, ready-to-use.
4. Substrate Solution (TMB): 15 mL, ready-to-use.
5. Stop Solution (0.5 M H₂SO₄): 15 mL, ready-to-use.
6. Extraction and sample dilution buffer (Carbonate buffer): 2 x 120 mL as 5x concentrate, dyed red. Dilute 1+4 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Plastic bag to store unused microtiter strips.
9. Instruction Manual.

Materials required but not provided

Instrumentation

1. 100 - 1000 µL micropipettes
2. Analytical balance
3. Mortar, mixer
4. Water bath
5. Centrifuge
6. ELISA reader (450 nm)

Reagents

Double-distilled water

ASSAY PROCEDURE

The washing solution is supplied as 10x concentrate and has to be diluted 1+9 with double distilled water before use.

In any case the diluted standards should be determined at least twofold. When samples in great numbers are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100µL ready-to-use standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual

liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.

5. Pipet 100µL of conjugate (anti-macadamia nut-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100µL of stop solution (0.5 M H₂SO₄) into each well. The blue color will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The color is stable for 30 minutes.

RESULTS

The diluted standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above-stated sample preparation procedure is already considered. Additional dilution due to meat-containing samples or high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppm on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of macadamia nut in ppm from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

For calculation of the amount of a corresponding raw product, the milk protein concentration has to be multiplied with a product-specific conversion factor (F).


The following conversion factors have been determined by means of validation experiments:

Non-fat milk powder (NIST RM1549)	2.7
Whole milk powder (NIST RM8435)	4.4
Caseinate	1.0
β-Lactoglobulin	1.1

Typical Standard Values

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 10 ppm standard. These values are only an example and should not be used instead of the standard.

Milk protein (ppm)	% binding of 10 ppm
10	100
6	86
1	50
0.4	32
0	4



PERFORMANCE CHARACTERISTICS

Sensitivity

The limit of detection (LOD) of the **DAI Milk ELISA** test is 0.05 ppm of milk protein.

Validation experiments with common matrices resulted in the following LODs (ppm).

Soy milk	0.13
Orange juice	1.10
White wine	0.03
Bread crumbs	0.08
Cookies	0.16
Chocolate	0.10
Sausage	0.18

The limit of quantification (LOQ) of the **DAI Milk ELISA** test is 0.4 ppm of milk protein.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

Specificity

For the following foods no cross-reactivity could be detected:

Egg	Sesame	Almond
Wheat	Mustard	Cocoa
Rye	Lupin	Beef
Barley	Celery	Pork
Oats	Peanut	Chicken
Rice	Hazelnut	Cod
Corn	Pistachio	
Soy	Walnut	

The following cross reactions were determined:

Ewe's milk	0.94%
Goat's milk	0.01%

Precision

Intra-assay Precision	8-10%
Inter-assay Precision	10-17%

Linearity

The serial dilution of spiked samples (cookies, bread crumbs, chocolate, sausage, soy milk, orange juice and white wine) resulted in a dilution linearity of 80 - 130%.

Recovery

Mean recovery was determined by spiking samples with different amounts of casein:

Cookies	102%
Bread crumbs	110%
Chocolate	99%
Sausage	88%
Soy milk	79%
Orange juice	106%
White wine	122%

PRECAUTIONS

Full compliance with the following good laboratory practices (GLP) will determine the reliability of the results:

1. Before beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipettes, ELISA reader etc.).

HEALTH & SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

REFERENCES


1. De Luis R, et al. (2007) – Development of two immunoassay formats to detect β -lactoglobulin. *J of Food Protection*, 70(7):1691-97
2. Restani P, et al. (1999) – Cross-reactivity between milk proteins from different animal species. *Clin Exp Allergy*, 29(7):997-1004
3. Mäkinen-Kiljunen S, et al. (1992) – A sensitive enzyme-linked immunosorbent assay for determination of bovine beta-lactoglobulin. *Allergy*, 47(4):347-52
4. Hefle SL, et al. (2004) – Validated sandwich enzyme-linked immunosorbent assay for casein and its application to retail milk-allergic complaint foods. *J Food Prot*, 67(9):1933-38




5. Patrick W, et al. (2009) – Determination of the bovine food allergen casein in white wines by quantitative indirect ELISA, SDS-Page, Western blot and immunostaining. *J Agric Food Chem*, 57(18):8399-405
6. Watanabe H et al. (2005) - Study on detection of allergenic substances (egg and milk) in processed meat products and frozen foods. *Sho Eis Zas*, 46(4):139-47
7. Downs ML, Taylor SL (2010) – Effects of thermal processing on the enzyme-linked immunosorbent assay (ELISA) detection of milk residues in a model food matrix. *J Agric Food Chem*, 58(18):10085-91
8. De Luis R, et al. (2007) – Development of two immunoassay formats to detect β -lactoglobulin: influence of heat treatment on β -lactoglobulin immunoreactivity. *J Food Prot*, 70(7):1691-7
9. Abbott M, et al. (2010) – Validation procedures for quantitative food allergen ELISA methods: community guidance and best practices. *J AOAC Int*, 93(2):442-50
10. Levin ME, et al. (2005) – Anaphylaxis in a milk-allergic child after ingestion of soy formula cross contaminated with cow's milk protein. *Pediatrics*, 116(5):1223-5

MANUFACTURER AND BRAND DETAILS

ISO 13485:2016



ISO 13485
Quality
Management for
Medical Devices
CERTIFIED

 **Diagnostic Automation/Cortez Diagnostics, Inc.**
21250 Califa Street, Suite 102 and 116,
Woodland Hills, California 91367 USA

Date Adopted	2024-03
Brand Name	AccuDiag™
REF 5165	AccuDiag™ - Milk ELISA
PIC	FD5165ZR44

Revision Date: 2017-03-31