



OneStep
hCG Combo Urine/Serum/Plasma
RapiDip™ InstaTest

REF 113030

PIC FT113030VV86

IVD

hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest	
Principle	Immunochromatographic Assay
Detection	Qualitative
Sensitivity	25 mIU/ml

PRODUCT FEATURES

- High Sensitivity and Specificity**
- Built in Internal Controls**
- Simplified Workflow, Non-invasive, and Safe**
- Rapid and Faster Results**

INTENDED USE

Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest is a rapid chromatographic immunoassay for the qualitative detection of human chorionic gonadotropin in urine or serum or plasma to aid in the early detection of pregnancy.

SIGNIFICANCE AND SUMMARY

Human chorionic gonadotropin (hCG) is a glycoprotein hormone produced by the developing placenta shortly after fertilization. In normal pregnancy, hCG can be detected in both urine and serum or plasma as early as 7 to 10 days after conception.^{1,2,3,4} hCG levels continue to rise very rapidly, frequently exceeding 100 mIU/ml by the first missed menstrual period,^{2,3,4} and peaking in the 100,000-200,000 mIU/ml range about 10-12 weeks into pregnancy. The appearance of hCG in both the urine and serum or plasma soon after conception, and its subsequent rapid rise in concentration during early gestational growth, make it an excellent marker for the early detection of pregnancy.

Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest is a rapid test that qualitatively detects the presence of hCG in urine or serum or plasma specimen at the sensitivity of 25mIU/ml. The test utilizes a combination of monoclonal and polyclonal antibodies to selectively detect elevated levels of hCG in urine or serum or plasma. At the level of claimed sensitivity, hCG Combo RapiDip shows no cross-reactivity interference from the structurally related glycoprotein hormones hFSH, hLH and hTSH at high physiological levels.

ASSAY PRINCIPLE

Cortez Diagnostics Inc. OneStep hCG Comb Urine/Serum/Plasma RapiDip™ InstaTest is a rapid chromatographic immunoassay for the qualitative detection of human chorionic gonadotropin in urine or serum or plasma to aid in the early detection of pregnancy. The test uses two lines to indicate results. The test utilizes a combination of antibodies including a monoclonal hCG antibody to selectively detect elevated levels of hCG. The control line is composed of goat polyclonal antibodies and colloidal gold particles. The assay is conducted by immersing the test dipstick in a urine or serum or plasma specimen and observing the formation of colored lines. The specimen migrates via capillary action along the membrane to react with the colored conjugate.

Positive specimens react with the specific antibody-hCG-colored conjugate to form a colored line at the test line region of the membrane. Absence of this colored line suggests a negative result. To serve as a procedural control, a colored line will always appear in the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

SPECIMEN COLLECTION & PREPARATION

Urine Assay

A urine specimen must be collected in a clean and dry container. A first morning urine specimen is preferred since it generally contains the highest concentration of hCG; however, urine specimens collected at any time of the day may be used. Urine specimens exhibiting visible precipitates should be centrifuged, filtered, or allowed to settle to obtain a clear specimen for testing.



Serum or plasma Assay

Blood should be collected aseptically into a clean tube without anticoagulants (Serum) or with anticoagulants (Plasma). Separate the serum or plasma from blood as soon as possible to avoid hemolysis. Use clear non-hemolyzed specimens when possible.

Specimen Storage

Urine or serum or plasma specimens may be stored at 2-8°C for up to 48 hours prior to testing. For prolonged storage, specimens may be frozen and stored below -20°C. Frozen specimens should be thawed and mixed before testing.

REAGENTS

Materials provided with the kit

- The test contains anti-hCG particles and anti-hCG coated on the membrane.
- Test dipstick, package insert

Materials required but not provided

- Specimen collection containers
- timer

ASSAY PROCEDURE

1. Bring the pouch or canister to room temperature before opening it. Remove the test dipstick from the sealed pouch or closed canister and use it within one hour.


Note: For canister packaging, immediately close the canister tightly after removing the required number of the test dipstick(s). Record the initial opening date on the canister. Once the canister has been opened, the remaining test dipstick(s) are stable for 90 days only.

2. With arrows pointing toward the urine or serum or plasma specimen, immerse the test dipstick vertically in the urine or serum or plasma specimen for at least 15

seconds. Do not pass the maximum line (MAX) on the test dipstick when immersing the dipstick. See illustration below.

3. Place the test dipstick on a non-absorbent flat surface, start the timer and wait for the colored line(s) to appear. **Read the result at 3 minutes when testing a urine specimen, or at 5 minutes when testing a serum or plasma specimen.**

Note: A low hCG concentration might result in a weak line appearing in the test line region (T) after an extended period of time; therefore, do not interpret the result after 10 minutes.



QUALITY CONTROL

A procedural control is included in the test. A colored line appearing in the control line region (C) is considered an internal procedural control. It confirms sufficient specimen volume and correct procedural technique. A clear background is an internal negative procedural control. If a background color appears in the result window and interferes with the ability to read the test result, the result may be invalid. It is recommended that a positive hCG control

(containing 25-250mIU/ml hCG) and a negative hCG control (containing "0"mIU/ml hCG) be evaluated to verify proper test performance when a new shipment of tests is received.

RESULTS

POSITIVE: Two distinct colored lines appear. One line should be in the control line region (C) and another line should be in the test line region (T). One line may be lighter than the other; they do not have to match. This means that you are probably pregnant.

NEGATIVE: One colored line appears in the control line region (C). No line appears in the test line region (T). This means that you are probably not pregnant.

INVALID: The result is invalid if no colored line appears in the control line region (C), even if a line appears in the test line region (T). You should repeat the test with a new test dipstick.

LIMITATIONS OF THE ASSAY

1. Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest is a preliminary qualitative test, therefore, neither the quantitative value nor the rate of increase in hCG can be determined by this test.
2. Very dilute urine specimens, as indicated by a low specific gravity, may not contain representative levels of hCG. If pregnancy is still suspected, a first morning urine specimen should be collected 48 hours later and tested.
3. Very low levels of hCG (less than 50 mIU/ml) are present in urine and serum or plasma specimen shortly after implantation. However, because a significant number of first trimester pregnancies terminate for natural reasons, a test result that is weakly positive should be confirmed by retesting with a first morning urine or

Diagnostic Automation/Cortez Diagnostics, Inc.

21250 Califa St, Suite 102 and 116, Woodland Hills, CA 91367 USA Phone: 818-591-3030, Fax: 818-591-8383

Email: onestep@rapidtest.com Website: www.rapidtest.com



- serum or plasma specimen collected 48 hours later.
- This test may produce false positive results. A number of conditions other than pregnancy, including trophoblastic disease and certain non-trophoblastic neoplasms including testicular tumors, prostate cancer, breast cancer, and lung cancer, cause elevated levels of hCG.^{6,7} Therefore, the presence of hCG in urine or serum or plasma specimens should not be used to diagnose pregnancy unless these conditions have been ruled out.
 - This test may produce false negative results. False negative results may occur when the levels of hCG are below the sensitivity level of the test. When pregnancy is still suspected, a first morning urine specimen should be collected 48 hours later and tested. In case pregnancy is suspected and the test continues to produce negative results, see a physician for further diagnosis.
 - As with any assay employing mouse antibodies, the possibility exists for interference by human anti-mouse antibodies (HAMA) in the specimen. Specimens from patients who have received preparations of monoclonal antibodies for diagnosis or therapy may contain HAMA. Such specimens may cause false positive or false negative results.
 - This test provides a presumptive diagnosis for pregnancy. A confirmed pregnancy diagnosis should only be made by a physician after all clinical and laboratory findings have been evaluated.

EXPECTED RANGE OF VALUES

Negative results are expected in healthy non-pregnant women and healthy men. Healthy pregnant women have hCG present in their urine and serum or plasma specimens. The amount of hCG will vary greatly with gestational age and between individuals. The hCG combo RapiDip™ (Urine/Serum/Plasma) has a sensitivity of 25 mIU/ml, and is

capable of detecting pregnancy as early as 1 day after the first missed menses.

PERFORMANCE CHARACTERISTICS

Accuracy

A multi-center clinical evaluation was conducted comparing Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest to another commercially available urine and serum or plasma hCG Rapid test. The urine study included 608 specimens, and both assays identified 377 negative and 231 positive results. The serum or plasma study included 308 specimens, and both assays identified 240 negative and 68 positive results. The results demonstrated a >99% overall accuracy of Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest when compared to the other urine and serum or plasma hCG Rapid test.

hCG Reference Method (Urine)

Method	Other hCG Rapid Test		Total Results
	Positive	Negative	
hCG Combo RapiDip™	231	0	231
	0	377	377
Total Results	231	377	608
Sensitivity:>99.9% (98.7%~100%)*		Specificity:>99.9% (99.2%~100%)*	
Accuracy: >99.9%(99.5%~100%)* 95% Confidence Intervals			

hCG Reference Method (Serum or Plasma)

Method	Other hCG Rapid Test		Total Results
	Positive	Negative	
hCG Combo RapiDip™	68	0	68
	0	240	240

Total Results	68	240	308
Sensitivity:>99.9% (95.7%~100%)*	Specificity:>99.9% (98.8%~100%)*		
Accuracy:>99.9% (99.0%~100%)*	* 95% Confidence Intervals		

Tests were performed with serum and found to be positive or negative as reported. Plasma specimens from the same individuals were tested with Cortez Diagnostics Inc. hCG (S/P/U) Tests to validate the efficacy with plasma specimens.

Sensitivity and Cross-Reactivity

Cortez Diagnostics Inc. OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest detects hCG at a concentration of 25 mIU/ml or greater.

The test has been standardized to the W.H.O. International Standard. The addition of LH (300mIU/ml), FSH (1,000mIU/ml), and TSH (1,000µIU/ml) to negative (0mIU/ml hCG) and positive (25mIU/ml hCG) specimens showed no cross-reactivity.

Precision

Intra-Assay

Within-run precision has been determined by using 10 replicates of three specimens containing 25 mIU/ml, 100 mIU/ml, 250 mIU/ml and 0 mIU/ml of HCG. The negative and positive values were correctly identified 100% of the time.

Inter-Assay

Between-run precision has been determined by using the same three specimens of 25 mIU/ml, 100 mIU/ml, 250 mIU/ml and 0 mIU/ml of HCG in 10 independent assays. Three different lots of the hCG combo RapidDip have been tested. The specimens were correctly identified 100% of the time.

Interfering Substance

The following potentially interfering substances were added to hCG negative and positive specimens.

Diagnostic Automation/Cortez Diagnostics, Inc.

21250 Califa St, Suite 102 and 116, Woodland Hills, CA 91367 USA Phone: 818-591-3030, Fax: 818-591-8383

Email: onestep@rapidtest.com Website: www.rapidtest.com



Acetaminophen	20 mg/dl	Caffeine	20 mg/dl
Acetylsalicylic Acid	20 mg/dl	Gentisic Acid	20 mg/dl
Ascorbic Acid	20 mg/dl	Glucose	2 g/dl
Atropine	20 mg/dl	Hemoglobin	1 mg/dl
Bilirubin	2 mg/dl	Bilirubin(serum or plasma)	40mg/dl
Triglycerides (serum or plasma)			1,200 mg/dl

None of the substances at the concentration tested interfered in the assay.

PRECAUTIONS

Please read all the information in this package insert before performing the test.

1. For professional in vitro diagnostic use only. Do not use after the expiration date.
2. The test should remain in the sealed pouch until ready to use.
3. All specimens should be considered potentially hazardous and handled in the same manner as an infectious agent.
4. The used test should be discarded according to local regulations.



STORAGE CONDITIONS

Store as packaged at room temperature or refrigerated (2-30°C). The test is stable through the expiration date printed on the sealed pouch or label of the closed canister. The test must remain in the sealed pouch or closed canister until use. **DO NOT FREEZE.** Do not use beyond the expiration date.

REFERENCES

1. Batzer FR. Hormonal evaluation of early pregnancy, Fertil. Steril. 1980; 34(1): 1-13
2. Catt KJ, ML Dufau, JL Vaitukaitis Appearance of hCG in pregnancy plasma following the initiation of implantation of the blastocyte, J. Clin. Endocrinol. Metab. 1975; 40(3): 537-540
3. Braunstein GD, J Razor, H. Danzer, D Adler, ME Wade Serum or plasma human chorionic gonadotropin levels throughout normal pregnancy, Am. J. Obstet. Gynecol. 1976; 126(6): 678-681
4. Lenton EA, LM Neal, R Sulaiman Plasma concentration of human chorionic gonadotropin from the time of implantation until the second week of pregnancy, Fertil. Steril. 1982; 37(6): 773-778
5. Steier JA, P Bergsjö, OL Myking Human chorionic gonadotropin in maternal plasma after induced abortion, spontaneous abortion and removed ectopic pregnancy, Obstet. Gynecol. 1984; 64(3): 391-394
6. Dawood MY, BB Saxena, R Landesman Human chorionic gonadotropin and its subunits in hydatidiform mole and choriocarcinoma, Obstet. Gynecol. 1977; 50(2): 172-181
7. Braunstein GD, JL Vaitukaitis, PP Carbone, GT Ross Ectopic production

MANUFACTURER AND BRAND DETAILS

 <p>ISO 13485:2016 Quality Management for Medical Devices CERTIFIED</p>	
 <p>Diagnostic Automation/Cortez Diagnostics, Inc. 21250 Califa Street, Suite 102 and 116, Woodland Hills, California 91367 USA</p>	
Date Adopted	2024-01
Brand Name	RapiDip™ InstaTest
REF 113030	OneStep hCG Combo Urine/Serum/Plasma RapiDip™ InstaTest
PIC	FT113030VV86
EU REP	AR Experts BV, Boeingavenue 201-219 1119 PD Schiphol-Rijk, The Netherlands info@ar-experts.eu
Revision Date: 2017-12-13	