



AccuDiag™ Mycoplasma pneumoniae IgM ELISA Kit

REF 8043-P1

IVD See External Label 2°C 96 Tests

threatening (6 - 8). *Mycoplasma pneumoniae*, a prokaryote, is the smallest (10 x 200nm), and simplest self-replicating microorganism known, and more closely resembles a bacterium rather than a virus. However, because it lacks a cell-wall, a resistance to cell-wall-active antibiotics is obvious (i.e., penicillin, cephalosporins (1)). This concern for diagnostic, or at least therapeutic accuracy in the early management of community-acquired infections is particularly critical in very young or elderly patients where very little temporal margin of error exists. Until recently, the routine laboratory diagnosis of this infection has been limited to insensitive and/or non-specific assays (i.e., cold agglutinins, complement-fixation, culture isolation). Research shows that species-specific antibodies to surface antigens exist. They are protective, and are readily detected by ELISA; even in the early stages of the disease. The diagnosis therefore, is best achieved serologically (9).

ASSAY PRINCIPLE

The DAI *Mycoplasma pneumoniae* IgM ELISA test system is designed to detect IgM class antibodies to *M. pneumoniae* in human sera. Wells of plastic micro well strips are sensitized by passive absorption with *M. pneumoniae* antigen. The test procedure involves three incubation steps:

1. Test sera are diluted with the Sample Diluent provided. The Sample Diluent contains anti-human IgG which precipitates and removes IgG and rheumatoid factor from the sample leaving IgM free to react with the immobilized antigen. During sample incubation, any antigen specific IgM antibody in the sample will bind to the immobilized antigen. The plate is washed to remove unbound antibody and other serum components.
2. Peroxidase Conjugated goat anti-human IgM is added to the wells and the plate is incubated. The Conjugate will react with IgM antibody immobilized on the solid phase in step 1. The wells are washed to remove unbound Conjugate.
3. The microwells containing immobilized peroxidase Conjugate are incubated with peroxidase Substrate Solution. Hydrolysis of the Substrate by peroxidase produces a color change. After a period of time the reaction is stopped and the color intensity of the solution is measured photometrically. The color intensity of the solution depends upon the antibody concentration in the original test sample.

SPECIMEN COLLECTION & PREPARATION

1. DAI recommends that the user carry out specimen collection in accordance with CLSI document M29: Protection of Laboratory Workers from Infectious Disease (Current Edition).
2. No known test method can offer complete assurance that human blood samples will not transmit infection. Therefore, consider all blood derivatives potentially infectious.
3. Use only freshly drawn and properly refrigerated sera obtained by approved aseptic venipuncture procedures in this assay (10, 11). Do not use if there are any added anticoagulants or preservatives. Avoid using hemolyzed, lipemic, or bacterially contaminated sera.
4. Store sample at room temperature for no longer than 8 hours. If testing is not performed within 8 hours, sera may be stored between 2 - 8°C, for no longer than 48 hours. If a delay in testing is anticipated, store test sera at -20°C or lower. Avoid multiple freeze/thaw cycles which may cause loss of antibody activity and give erroneous results. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine stability criteria for its laboratory (13).

Mycoplasma pneumoniae IgM ELISA	
Principle	Indirect ELISA
Detection	Qualitative
Sample	10 µL serum/plasma
Incubation Time	60 minutes
Sensitivity	100%
Specificity	93%
Shelf Life	12 Months from the manufacturing date

PRODUCT FEATURES

- Very easy to use with little training
- Highly specific and consistent Assay
- Provides accurate results quickly
- Reading of results both visually and as absorbance data

INTENDED USE

The Diagnostics Automation Inc.(DAI) *Mycoplasma pneumoniae* IgM Test System provides a means for the qualitative detection of IgM antibodies to *Mycoplasma pneumoniae* in human sera. When performed according to these instructions, the results of this test may aid in the diagnosis of *M. pneumoniae* infections in the adult population. This assay is for *In Vitro* diagnostic use only.

SIGNIFICANCE AND SUMMARY

Mycoplasma pneumoniae is the most common cause of pneumonia and febrile upper-respiratory tract infections in the general population (except for influenza A) (1 - 5). Other nonrespiratory complications may also develop with this disease in virtually any organ system, with insult ranging from mild to life-



REAGENTS

Materials provided with the kit

Each kit contains the following components in sufficient quantities to perform the number of tests indicated on packaging label.

Note: All reactive reagents contain sodium azide as a preservative at a concentration of 0.1% (w/v): Controls, Calibrator and Sample Diluent.

1. Plate: 96 wells configured in twelve 1x8-well strips coated with inactivated preparation of M. pneumonia (strain FH) antigen. The strips are packaged in a strip holder and sealed in an envelope with desiccant.
2. Conjugate: Conjugated (horseradish peroxidase) goat anti-human IgM (μ chain specific). Ready to use. One, 15 mL vial with a white cap.
3. Positive Control (Human Serum): One, 0.35 mL vial with a red cap.
4. Calibrator (Human Serum): One, 0.5 mL a blue cap.
5. Negative Control (Human Serum): One, 0.35 mL vial with a green cap.
6. Sample Diluent: One 30 mL bottle (green cap) containing Tween-20, bovine serum albumin and phosphate- buffered-saline and goat anti-human IgG (γ -chain specific). Purple solution. Ready to use.
7. TMB: One 15 mL amber bottle (amber cap) containing 3,3',5,5'-tetramethylbenzidine (TMB). Ready to use.
8. Stop solution: One 15 mL bottle (red cap) containing 1M H₂SO₄, 0.7MHCl. Ready to use.
9. Wash buffer concentrate (10X): dilute 1 part concentrate + 9 parts deionized or distilled water. One 100 mL bottle (clear cap) containing a 10X concentrated phosphate-buffered-saline and Tween-20 solution (blue solution). **NOTE: 1X solution will have a pH of 7.2 \pm 0.2.**

NOTES:

1. The following components are not kit lot number dependent and may be used interchangeably with the ELISA assays: TMB, Stop Solution, and Wash Buffer.
2. Component list containing lot specific information is inside the kit box.

Materials required but not provided

1. ELISA microwell reader capable of reading at a wavelength of 450nm. **NOTE: Use of a single (450nm), or dual (450/620 - 650nm), wavelength reader is acceptable. Dual wavelength is preferred, as the additional reference filter has been determined to reduce potential interference from anomalies that may absorb light.**
2. Pipettes capable of accurately delivering 10 - 200 μ L.
3. Multichannel pipette capable of accurately delivering (50 - 200 μ L).
4. Reagent reservoirs for multichannel pipettes.
5. Wash bottle or microwell washing system.
6. Distilled or deionized water.
7. One liter graduated cylinder.
8. Serological pipettes.
9. Disposable pipette tips.
10. Paper towels.
11. Laboratory timer to monitor incubation steps.
12. Disposal basin and disinfectant. (example: 10% household bleach, 0.5% sodium hypochlorite.)

ASSAY PROCEDURE

1. Remove the individual components from storage and allow them to warm to room temperature (20-25°C).
2. Determine the number of microwells needed. Allow six Control/Calibrator determinations (one Blank, one Negative Control, three Calibrators and one Positive Control) per run. A Reagent Blank should be run on each assay. Check software and reader requirements

for the correct Controls/ Calibrator configurations. Return unused strips to the resealable pouch with desiccant, seal, and return to storage between 2-8°C.

EXAMPLE PLATE SET-UP

	1	2
A	Blank	Patient 3
B	Neg. Control	Patient 4
C	Calibrator	Etc.
D	Calibrator	
E	Calibrator	
F	Pos. Control	
G	Patient 1	
H	Patient 2	

3. Prepare a 1:21 dilution (e.g.: 10 μ L of serum + 200 μ L of Sample Diluent. of the Negative Control, Calibrator, Positive Control, and each patient serum.
4. To individual wells, add 100 μ L of each diluted control, calibrator and sample. Ensure that the samples are properly mixed. Use a different pipette tip for each sample.
5. Add 100 μ L of Sample Diluent to well A1 as a reagent blank. Check software and reader requirements for the correct reagent blank well configuration.
6. Incubate the plate at room temperature (20-25°C) for 25 \pm 5 minutes.
7. Wash the microwell strips 5 times.
 - a. **Manual Wash Procedure:**
 1. Vigorously shake out the liquid from the wells.
 2. Fill each micro well with Wash Buffer. Make sure no air bubbles are trapped in the wells.
 3. Repeat steps a. and b. for a total of 5 washes.
 4. Shake out the wash solution from all the wells. Invert the plate over a paper towel and tap firmly to remove any residual wash solution from the wells. Visually inspect the plate to ensure that no residual wash solution remains. Collect wash solution in a disposable basin and treat with disinfectant at the end of the day's run.
 - b. **Automated Wash Procedure:**
If using an automated microwell wash system, set the dispensing volume to 300-350 μ L/well. Set the wash cycle for 5 washes with no delay between washes. If necessary, the microwell plate may be removed from the washer, inverted over a paper towel and tapped firmly to remove any residual wash solution from the microwells.
8. Add 100 μ L of the Conjugate to each well, including reagent blank well, at the same rate and in the same order as the specimens were added.
9. Incubate the plate at room temperature (20-25°C) for 25 \pm 5 minutes.
10. Wash the microwells by following the procedure as described in step 7.
11. Add 100 μ L of TMB to each well, including reagent blank well, at the same rate and in the same order as the specimens were added.
12. Incubate the plate at room temperature (20-25°C) for 10 to 15 minutes.
13. Stop the reaction by adding 50 μ L of Stop Solution to each well, including reagent blank well, at the same rate and in the same order as the TMB was added. Positive samples will turn from blue to yellow. After adding the Stop Solution, tap the plate several times to ensure that the samples are thoroughly mixed.
14. Set the microwell reader to read at a wavelength of 450nm and measure the optical density (OD) of each well against the reagent blank. The plate should be read within 30 minutes after the addition of the Stop Solution.



ABBREVIATED TEST PROCEDURE

1. Dilute Serum 1:21.
2. Add diluted sample to microwell – 100µL/well.
3. $\xrightarrow{\hspace{2cm}}$ Incubate \pm 5 minutes.
4. Wash.
5. Add Conjugate – 100µL/well.
6. $\xrightarrow{\hspace{2cm}}$ Incubate \pm 5 minutes.
7. Wash.
8. Add TMB – 100µL/well.
9. $\xrightarrow{\hspace{2cm}}$ Incubate 10 - 15 minutes.
10. Add Stop Solution – 50µl/well – Mix.
11. READ within 30 minutes.

RESULTS

Calculations:

- a. **Correction Factor:** The manufacturer determined a Cutoff OD Value for positive samples and correlated it to the Calibrator. The Correction Factor (CF) allows for the determination of the Cutoff Value for positive samples. It will also correct for slight day-to-day variations in test results. The Correction Factor is determined for each lot of components and is printed on the Component Label located in the Test System box.
- b. **Cutoff OD Value:** To obtain the Cutoff OD Value, multiply the CF by the mean OD of the Calibrator determined above. (CF x Mean OD of Calibrator = Cutoff OD Value)
- c. **Index Values/OD Ratios:** Calculate the Index Value/OD Ratio for each specimen by dividing its OD Value by the Cutoff OD from step b.

Example:

Mean OD of Calibrator	0.793
Correction Factor (CF)	0.25
Cut off OD	$0.793 \times 0.25 = 0.198$
Unknown Specimen OD	0.432
Specimen Index Value or OD Ratio	$0.432 / 0.198 = 2.18$

INTERPRETATION

Interpretations: Index Values or OD ratios are interpreted as follows:

	Index Value or OD Ratio
Negative Specimens	≤ 0.90
Equivocal Specimens	0.91 to 1.09
Positive Specimens	≥ 1.10

- a. An OD ratio < 0.90 indicates no significant amount of antibodies to *M. pneumoniae* detected. A non-reactive result indicates no current/previous infection.
- b. An OD ratio > 1.10 indicates that IgM antibodies specific to *M. pneumoniae* were detected. A reactive test result indicates a past/recent infection.
- c. Specimens with OD ratio values in the equivocal range (0.91 - 1.09) should be retested in duplicate. Report any two of the three results which agree. Evaluate repeatedly equivocal using an alternate serological method and/or re-evaluate by drawing another sample one to three weeks later.

Note: The magnitude of the measured result above the cut-off is not indicative of the total amount of antibody present and cannot be correlated to IFA titers.

QUALITY CONTROL


1. Each time the assay is performed, the Calibrator must be run in triplicate. A reagent blank, Negative Control, and Positive Control must also be included in each assay.
2. Calculate the mean of the three Calibrator wells. If any of the three values differ by more than 15% from the mean, discard that value and calculate the mean using the remaining two wells.
3. The mean OD value for the Calibrator and the OD values for the Positive and Negative Controls should fall within the following ranges:

	OD Range
Negative Control	≤ 0.250
Calibrator	≥ 0.300
Positive Control	≥ 0.500

- a. The OD of the Negative Control divided by the mean OD of the Calibrator should be ≤ 0.9 .
 - b. The OD of the Positive Control divided by the mean OD of the Calibrator should be ≥ 1.25 .
 - c. If the above conditions are not met the test should be considered invalid and should be repeated.
4. The Positive Control and Negative Control are intended to monitor for substantial reagent failure and will not ensure precision at the assay Cutoff.
 5. Additional controls may be tested according to guidelines or requirements of local, state, and/or federal regulations or accrediting organizations.
 6. Refer to NCCLS document C24: Statistical Quality Control for Quantitative Measurements for guidance on appropriate QC practices.

EXPECTED RANGES OF VALUES

The clinical study for this product included 220 random specimens sent to a reference laboratory in the northeastern United States for routine *Mycoplasma* serological analysis. With respect to this population, 201/220 (91.4%) were negative, 3/220(1.4%) were equivocal, and 16/220 (7.3%) were reactive. In addition, an in-house study evaluated 180 random normal donor sera. Depiction of results follows in the frequency distribution chart.



PERFORMANCE CHARACTERISTICS

Comparative Study

A comparative study was performed to demonstrate the equivalence of the DAI *Mycoplasma* IgM ELISA test system. Technicians evaluated the performance of the DAI ELISA *M. pneumoniae* IgM Test System of 299



specimens, in a three-site clinical investigation. All clinical sites compared the performance of the DAI ELISA to the IFA test system. Table 1 shows a summary of the testing performed at each clinical site. Table 2 shows the results of this comparative testing.

Table 1: Summary of Clinical Testing

Site	Location	Specimen Characteristics	n
One	Offsite	Routine specimens which were sent to a reference laboratory in Northeastern U.S. for Mycoplasma serological analysis	111
One	Offsite	Samples sent to a hospital in the Midwest for Mycoplasma serological analysis.	9
Two	Offsite	Routine specimens which were sent to a reference laboratory in Northeastern U.S. for Mycoplasma serological analysis.	100
Two	Offsite	Repository specimens previously tested for Mycoplasma IgM and were found to be reactive.	2
Three	In-house	Various disease-state paired sera from diagnosed Mycoplasma infections.	62
Three	In-house	Disease-state specimens from confirmed Mycoplasma infections.	15

Table 2: Clinical Site One - Calculation of Relative Sensitivity, Specificity and Agreement

IFA Results	DAI Mycoplasma IgM ELISA			
	Negative	Equivocal	Positive	Total
<1:16	102	1	0	103
1:16	8	0	0	8
≥ 1:32	2	2	5	9
Totals	112	3	5	120

Relative Sensitivity = 5/7 = 71.4% 95% Confidence Interval * = 29.0 to 96.3%
 Relative Specificity = 102/102 = 100% 95% Confidence Interval * = 96.4 to 100%
 Relative Agreement = 107/109 = 98.2% 95% Confidence Interval * = 93.5 to 99.8%
 *95% confidence intervals calculated using the exact method

Table 3: Clinical Site Two - Calculation of Relative Sensitivity, and Agreement

IFA Results	DAI Mycoplasma IgM ELISA			
	Negative	Equivocal	Positive	Total
<1:16	89	0	7	96
1:16	0	0	0	0
≥ 1:32	0	0	6	6
Totals	89	0	13	102

Relative Sensitivity = 6/6 = 100% 95% Confidence Interval * = 54.1 to 100%
 Relative Specificity = 89/96 = 92.7% 95% Confidence Interval * = 85.6 to 97.0%
 Relative Agreement = 95/102 = 93.1% 95% Confidence Interval * = 86.4 to 97.2%
 *95% confidence intervals calculated using the exact method

Table 4: Clinical Site Three - Calculation of Relative Sensitivity, and Agreement

IFA Results	DAI Mycoplasma IgM ELISA			
	Negative	Equivocal	Positive	Total
<1:16	27	1	10	38
1:16	0	0	5	5
≥ 1:32	3	1	30	34
Totals	30	2	45	77

Relative Sensitivity = 30/33 = 90.9% 95% Confidence Interval * = 75.7 to 98.1%
 Relative Specificity = 27/37 = 73.0% 95% Confidence Interval * = 55.9 to 86.2%
 Relative Agreement = 57/70 = 81.4% 95% Confidence Interval * = 70.3 to 89.7%

*95% confidence intervals calculated using the exact method

Table 5: All Sites Combined - Calculation of Relative Sensitivity, and Agreement

IFA Results	DAI Mycoplasma IgM ELISA			
	Negative	Equivocal	Positive	Total
<1:16	218	2	17	237
1:16	8	0	5	13
≥ 1:32	5	3	41	49
Totals	231	5	63	299

Relative Sensitivity = 41/46 = 89.1% 95% Confidence Interval * = 76.4 to 96.4%
 Relative Specificity = 218/235 = 92.8% 95% Confidence Interval * = 88.7 to 95.7%
 Relative Agreement = 259/281 = 92.2% 95% Confidence Interval * = 88.4 to 95.0%
 *95% confidence intervals calculated using the exact method

Note: Be advised that relative refers to the comparison of this assay's results to that of a similar assay. There was not an attempt to correlate the assay's results with the presence of absence of disease. Make no judgment on the comparison assay's accuracy to predict disease.

2. Precision and Reproducibility

Two clinical sites conducted reproducibility studies using the same eight specimens: two relatively strong positive specimens, two specimens near the cut off, two that were clearly negative and the kit's positive and negative controls. On each day of testing, the technician tested each of the eight specimens in triplicate. The clinical sites conducted this reproducibility study for a three-day period. Reproducibility was evaluated as outlined in the FDA guidance document; Review Criteria for In Vitro Diagnostic Devices for Detection of IgM Antibodies to Viral Antigens. A summary of this investigation appears in Tables 6 and 7 below.

Table 6: Summary of Intra-Assay Precision Testing Conducted at Clinical Sites One and Two

Sample (N)	Results at Site One						Results at Site Two					
	Day #1		Day #2		Day #3		Day #1		Day #2		Day #3	
	Mean Ratio	Percent CV	Mean Ratio	Percent CV	Mean Ratio	Percent CV	Mean Ratio	Percent CV	Mean Ratio	Percent CV	Mean Ratio	Percent CV
1	2.53	9.7	2.70	8.2	2.95	8.0	2.05	38.0	2.73	15.7	2.46	11.2
2	1.14	13.8	1.09	9.1	1.35	18.5	1.13	7.1	1.25	6.3	1.29	12.9
3	2.42	11.9	2.37	1.3	2.29	6.0	2.49	7.5	3.07	19.5	2.52	4.1
4	1.10	10.6	1.09	5.9	1.04	6.1	0.97	7.5	1.36	21.0	1.13	9.2
5	0.18	23.6	0.18	9.3	0.12	15.7	0.17	15.6	.013	50.0	0.19	8.2
6	0.20	23.1	0.24	5.2	0.17	8.5	0.16	16.1	0.18	14.7	0.23	13.5
NC	0.07	28.5	0.09	15.4	0.09	55.9	0.11	32.8	0.09	22.3	0.11	18.2
HPC	3.25	3.9	3.05	4.4	3.23	5.9	2.98	3.5	3.49	6.2	3.68	7.3

Table 7: Summary of Inter-Assay Precision Testing Conducted at Clinical Sites One and Two

Sample	Site One - Three Day Results		Site Two - Three Day Results	
	Mean Ratio	Percent CV	Mean Ratio	Percent CV
1	2.73	10.0	2.41	22.9
2	1.19	16.6	1.22	10.3
3	2.36	7.2	2.69	15.8
4	1.08	7.3	1.15	20.0
5	0.16	25.2	0.16	27.7
6	0.20	18.9	0.19	20.8
NC	0.09	36.2	0.10	23.5
HPC	3.18	5.1	3.38	10.6



Note: The reproducibility results depicted above are presented only as an example of those results obtained during the clinical study, using ideal conditions of environment, equipment and technique. Evaluate reproducibility at each laboratory, and may vary, depending upon the conditions at the laboratory.

LIMITATIONS OF THE ASSAY

1. Do not make a diagnosis based on ELISA *M. pneumoniae* IgM Test System results alone. Interpret test results in conjunction with clinical evaluation and results of other diagnostic procedures.
2. If testing a particular specimen occurs early during the primary infection, no detectable IgM may be evident. If there is suspicion of a Mycoplasma infection, take a second sample at least fourteen days later for additional testing.
3. A non-reactive result does not rule out current *M. pneumoniae* infection since the specimen may have been collected before demonstrable antibody was present or after the antibody has decreased below detectable levels. Consequently, demonstration of elevated IgG titers, in conjunction with specific IgM, increases the specificity of serological diagnosis.
4. Avoid the use of hemolytic, lipemic, bacterially contaminated, or heat-inactivated specimens. Erroneous results may occur.
5. DAI has not established assay performance characteristics for matrices other than serum.
6. DAI did not conduct Cross Reactivity Studies of the performance of this assay with certain types of specimens. These specimens include the following: those known to be positive for antibodies to organisms known to be associated with lower respiratory illness (i.e., Influenza A and B, CMV, *C. pneumoniae*, parainfluenza), those closely related Mycoplasma serovars known to cross-react with *M. pneumoniae*, such as *M. genitalium* and *M. hominis*, as well as various Ureaplasma species.
7. Do not use Mycoplasma culture results, or the presence or absence of antibody, to determine the success or failure of therapy.
8. Interpret specimens from immunocompromised patients with caution.
9. Do not perform screening of the general population. Test only when clinical characteristics are present or exposure is expected.
10. Studies show that the IgG removal system included with this test system will functionally remove the IgG from specimens containing total IgG levels ranging from 300 to 600 mg/mL. Studies were not conducted to establish the effectiveness of this removal system at IgG levels exceeding 600 mg/mL.
11. The prevalence of Mycoplasma IgM antibody is relatively low. Low-level prevalence rates of such analytes will affect the assay's predictive value.

PRECAUTIONS

1. For *In Vitro* diagnostic use.
2. Follow normal precautions exercised in handling laboratory reagents. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing, gloves, and eye/face protection. Do not breathe vapor. Dispose of waste observing all local, state, and federal laws.
3. The wells of the ELISA plate do not contain viable organisms. However, consider the strips **potentially biohazardous materials** and handle accordingly.
4. The Controls are **potentially biohazardous materials**. Source materials from which these products were derived were found negative for HIV-1 antigen, HBsAg and for antibodies against HCV and HIV by approved test methods. However, since no test method can offer complete assurance that infectious agents are absent, handle these products at the Biosafety Level 2 as recommended for any potentially infectious human serum or blood specimen in the Centers for Disease Control/National Institutes of Health manual "Biosafety in Microbiological and Biomedical Laboratories": Current Edition; and OSHA's Standard for Bloodborne Pathogens (12).
5. Adherence to the specified time and temperature of incubations is essential for accurate results. **All reagents must be allowed to reach room temperature (20 - 25°C) before starting the assay.** Return unused reagents to refrigerated temperature immediately after use.
6. Improper washing could cause false positive or false negative results. Be sure to minimize the amount of any residual wash solution; (e.g., by blotting or aspiration) before adding Conjugate or Substrate. Do not allow the wells to dry out between incubations.
7. The Sample Diluent, Controls, Conjugate and Wash Buffer contain Sodium Azide at a concentration of <0.1% (w/v). Sodium Azide has been reported to form lead or copper azides in laboratory plumbing which may cause explosions upon hammering. To prevent, rinse sink thoroughly with water after disposing of solution containing Sodium Azide.
8. The Stop Solution is TOXIC if inhaled, has contact with skin or if swallowed. It can cause burns. In case of accident or ill feelings, seek medical advice immediately.
9. The TMB Solution is HARMFUL. It is irritating to eyes, respiratory system and skin.
10. The Wash Buffer concentrate is an IRRITANT. It is irritating to eyes, respiratory system and skin.
11. Wipe the bottom of the plate free of residual liquid and/or fingerprints that can alter optical density (OD) readings.
12. Dilution or adulteration of these reagents may generate erroneous results.
13. Do not use reagents from other sources or manufacturers.
14. TMB Solution should be colorless, very pale yellow, very pale green, or very pale blue when used. Contamination of the TMB with Conjugate or other oxidants will cause the solution to change color prematurely. Do not use the TMB if it is noticeably blue in color.
15. Never pipette by mouth. Avoid contact of reagents and patient specimens with skin and mucous membranes.
16. Avoid microbial contamination of reagents. Incorrect results may occur.
17. Cross contamination of reagents and/or samples could cause erroneous results.
18. Reusable glassware must be washed and thoroughly rinsed free of all detergents.
19. Avoid splashing or generation of aerosols.
20. Do not expose reagents to strong light during storage or incubation.

STORAGE CONDITIONS

	Coated Microwell Strips: Immediately reseal extra strips with desiccant and return to proper storage. After opening - strips are stable for 60 days, as long as the indicator strips on the desiccant pouch remains blue.
	Conjugate – DO NOT FREEZE.
	Unopened Test System, Calibrator, Positive Control, Negative Control, TMB, Sample Diluent
	Stop Solution: 2 - 25°C
	Wash Buffer (1X): 20 - 25°C for up to 7 days, 2 - 8°C for 30 days.
	Wash Buffer (10X): 2 - 25°C



21. Allowing the microwell strips and holder to equilibrate to room temperature prior to opening the protective envelope will protect the wells from condensation.
22. Collect the wash solution in a disposal basin. Treat the waste solution with disinfectant (i.e.: 10% household bleach - 0.5% Sodium Hypochlorite). Avoid exposure of reagents to bleach fumes.
23. Caution: Neutralize any liquid waste at an acidic pH before adding to a bleach solution.
24. Do not use ELISA plate if the indicator strip on the desiccant pouch has turned from blue to pink.
25. Do not allow the Conjugate to come in contact with containers or instruments that may have previously contained a solution utilizing Sodium Azide as a preservative. Residual amounts of Sodium Azide may destroy the Conjugate's enzymatic activity.
26. Do not expose any of the reactive reagents to bleach-containing solutions or to any strong odors from bleach-containing solutions. Trace amounts of bleach (sodium hypochlorite) may destroy the biological activity of many of the reactive reagents within this Test System.

REFERENCES

1. Tazon CU, and Murray HW: "Atypical pneumonias". In: Respiratory Infections: diagnosis and Management. Pennington JE, ed. Raven Press, New York, NY, pp. 251, 1983.
2. Chanock RM, Fox HH, James WD, Gutekunst RR, White RT, Seterfit LB: Epidemiology of M.P. infection in military recruits. Ann. NY Acad. Sci. 143:484-496, 1967. DAI ELISA M. pneumoniae IgM Test System 6 (Rev. Date 10/13/2014)
3. Lind K, Bentzon MW: Epidemics of M. pneumoniae infection in Denmark from 1958 - 1974. Tnt. J. Epidemiol. 5:267-277, 1976.
4. Noah ND: M. pneumoniae infection in the United Kingdom. British Med. J. 2:544-546, 1974.
5. Foy HM, Kenny GE, Cooney MK, Allan ID: Long-term epidemiology of infections with M. pneumoniae. J. Infect. Dis. 139:681-687, 1979.
6. Murray HW, Masur H, Seterfit LB, and Roberts LB: The protean manifestation of M. pneumoniae infections in adults. Am. J. Med. 58:229-242, 1975.
7. Cassell GH, and Cole BC: Mycoplasmas as agents of human disease. N. Engl. J. Med. 304:80, 1981.
8. Noriega ER, Simberkoff MS, Gilroy SJ, et al: Life threatening M. pneumoniae. JAMA 29:1471-1472, 1974.
9. Carter JB, and Carter SC: Acute-phase, Indirect Fluorescent antibody Procedure for diagnosis of Mycoplasma pneumoniae infection. Ann. Clin. Lab. Sci. 13, No. 2, 150-155, 1983.
10. Procedures for the collection of diagnostic blood specimens by venipuncture - Second Edition: Approved Standard (1984). Published by National Committee for Clinical laboratory Standards.
11. Procedures for the Handling and Processing of Blood Specimens. NCCLS Document H18-A, Vol. 10, No. 12, Approved Guideline, 1990.
12. U.S. Department of Labor, Occupational Safety and Health Administration: Occupational Exposure to Bloodborne Pathogens, Final Rule. Fed. Register 56:64175- 64182, 1991.
13. Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guidelines – 4th Edition (2010). CLSI Document GP44- A4 (ISBN 1-56238-724-3). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, PA 19087.

MANUFACTURER AND BRAND DETAILS

ISO 13485:2016

ISO 13485
Quality
Management for
Medical Devices
CERTIFIED

Diagnostic Automation/Cortez Diagnostics, Inc.
21250 Califa Street, Suite 102 and 116,
Woodland Hills, California 91367 USA

Date Adopted	2023-11
Brand Name	AccuDiag™
REF 8043-P1	AccuDiag™ - Mycoplasma pneumoniae IgM ELISA
EC REP	CEpartner4U, Esdoornlaan 13, 3951 DB Maarn, The Netherlands www.cepartner4u.eu

Revision Date: 2017-12-20