



AccuDiag™ Rubella IgG ELISA Kit

REF 1301-P3

IVD See External Label 2°C 96 Tests

Rubella IgG ELISA	
Principle	Indirect ELISA
Detection	Qualitative
Sample	10 µL serum/plasma
Incubation Time	50 minutes
Shelf Life	12 Months from the manufacturing date

PRODUCT FEATURES

- Very easy to use with little training
- Highly specific and consistent Assay
- Provides accurate results quickly
- Reading of results both visually and as absorbance data

INTENDED USE

The Diagnostic Automation Inc. Rubella IgG ELISA Kit is intended for the detection of IgG antibody to Rubella in human serum or plasma.

SIGNIFICANCE AND SUMMARY

Rubella is usually a mild disease with infrequent complication. In unvaccinated populations, rubella is primarily a childhood disease. Where children are well-immunized, adolescent and adult infections become more evident. Rubella is spread by direct contact with nasal or throat secretions of infected individuals. Symptoms may include a rash, slight fever, joint aches, headache, discomfort, runny nose and reddened eyes. The incubation period for rubella is 12-23 days; in most cases, symptoms appear within 16-18 days. If contracted during the first trimester of pregnancy, Rubella infection can lead to congenital rubella syndrome (CRS). Infection of a pregnant woman may result in a miscarriage, stillbirth or the birth of an infant with abnormalities, which may include

deafness, cataracts, heart defects, liver and spleen damage and mental retardation. CRS occurs among at least 25 percent of infants born to women who have had rubella during the first trimester of pregnancy. The presence of IgG antibody to rubella virus is indicative of vaccination or previous exposure. In individuals with acute rubella infection, four-fold or greater increase in IgG antibody level is indicative of recent infection. Rubella IgM antibodies are detected by ELISA in 100% of patients between days 11-25 after onset of the exanthema, in 60-80% of individuals at days 15-25 after vaccination and in 90-97% of infants with congenital rubella between 2 weeks and 3 months after birth. Rubella IgM antibody often persists for 20-30 days after acute infection or vaccination.

ASSAY PRINCIPLE

Diluted patient serum is added to wells coated with purified antigen. IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

SPECIMEN COLLECTION AND HANDLING

1. Collect blood specimens and separate the serum.
2. Specimens may be refrigerated at 2-8°C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing.

REAGENTS

MATERIALS PROVIDED		96 Tests
1.	Microwells coated with Rubella antigen	12x8x1
2.	Sample Diluent: 1 Bottle (ready to use)	22 ml
3.	Calibrator: 1 Vial (ready to use)	1 ml
4.	Positive Control: 1 Vial (ready to use)	1 ml
5.	Negative Control: 1 Vial (ready to use)	1 ml
6.	Enzyme Conjugate: 1 Bottle (ready to use)	12 ml
7.	TMB Substrate: 1 Bottle (ready to use)	12 ml
8.	Stop solution: 1 Bottle (ready to use)	12 ml
9.	Wash Concentrate 20X: 1 Bottle	25 ml

Materials required but not provided

1. Distilled or deionized water
2. Precision pipettes
3. Disposable pipette tips
4. ELISA reader capable of reading absorbance at 450nm
5. Absorbance paper or paper towel
6. Graph paper

REAGENT PREPARATION

Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (18-26°C).



ASSAY PROCEDURE

Bring all specimens and kit reagents to room temperature (18-26°C) and gently mix.

1. Place the desired number of coated strips into the holder.
2. Negative control, positive control, and calibrator are ready to use. Prepare 1:21 dilution of test samples, by adding 10 µl of the sample to 200 µl of sample diluent. Mix well.
3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100 µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
4. Remove liquid from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20 minutes at room temperature.
6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel
7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature.
8. Add 100 µl of stop solution.
9. Read O.D. at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm.

CALCULATION OF RESULTS

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate the cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the O.D. value of each sample by cut-off value.

Example of typical results:

Calibrator mean OD = 0.8
 Calibrator Factor (CF) = 0.5
 Cut-off Value = 0.8 x 0.5 = 0.400
 Positive control O.D. = 1.2
 Ab Index = 1.2 / 0.4 = 3
 Patient sample O.D. = 1.6
 Ab Index = 1.6 / 0.4 = 4.0

QUALITY CONTROL

The test run may be considered valid provided the following criteria are met:

1. The O.D. of the Calibrator should be greater than 0.250.
2. The Ab index for Negative control should be less than 0.9.
3. The Ab Index for Positive control should fall within the range specified on the COA/label.

INTERPRETATION

The following is intended as a guide to interpretation of Rubella IgG test results; each laboratory is encouraged to establish its own criteria for test interpretation based on sample populations encountered.

Antibody Index Interpretation

<0.9 No detectable antibody to Rubella IgG by ELISA.

- 0.9-1.1 Borderline positive. Follow-up testing is recommended if clinically indicated.
- >1.1 Detectable antibody to Rubella IgG by ELISA.

LIMITATIONS OF THE ASSAY

1. The test results obtained using this kit serve only as an aid to diagnosis and should be interpreted in relation to the patient's history, physical findings and other diagnostic procedures.
2. Lipemic or hemolyzed samples may cause erroneous results.

STORAGE AND STABILITY

1. Store the kit at 2-8° C.
2. Keep microwells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light.

WARNINGS AND PRECAUTIONS

1. Potential biohazardous materials: The calibrator and controls contain human source components which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent. These reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories" 1984.
2. Optimal results will be obtained by strict adherence to the test protocol. Precise pipetting as well as following the exact time and temperature requirements is essential.
3. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
4. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
5. Control sera and sample diluent contain preserved with sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azide. On disposal, flush with a large volume of water.

REFERENCES

1. de Souza VA; Sumita LM; Otsubo ME; Takei K; Pannuti CS. Enzyme linked immunosorbent assay for rubella antibodies: a simple method of antigen production. A preliminary report. Rev Inst Med Trop Sao Paulo 1995; 37(4):357-9.
2. Matter L; Germann D; Bally F; Schopfer K. Age-stratified seroprevalence of Rubella, mumps and rubella (MMR) virus infections in Switzerland after the introduction of MMR mass vaccination. Eur J Epidemiol 1997;13(1):61-6.
3. M'uhlebach-Sponer M; Zbinden R; da Silva VA; Gnehm HE. Intrathecal rubella antibodies in an adolescent with Guillain-Barr'e syndrome after mumps-Rubella-rubella vaccination [letter]. Eur J Pediatr 1995; 154(2):166.
4. Johnson CE; Kumar ML; Whitwell JK; Staehle BO; Rome LP; Dinakar C; Hurni W; Nalin DR. Antibody persistence after primary Rubella-mumps-rubella vaccine and response to a second dose given at four to six vs. eleven to thirteen years. Pediatr Infect Dis J 1996; 15(8):687-92.




5. Matter L; Kogelschatz K; Germann D. Serum levels of rubella virus antibodies indicating immunity: response to vaccination of subjects with low or undetectable antibody concentrations. J Infect Dis 1997; 175(4):749-55.
6. Bos P; Steele D; Alexander J. Prevalence of antibodies to rubella, herpes simplex 2 and cytomegalovirus in pregnant women and in neonates at Ga-Rankuwa. Cent Afr J Med 1995; 41(1):14-7.

MANUFACTURER AND BRAND DETAILS

ISO 13485:2016



 Diagnostic Automation/Cortez Diagnostics, Inc.
 21250 Califa Street, Suite 102 and 116,
 Woodland Hills, California 91367 USA

Date Adopted	2023-11
Brand Name	AccuDiag™
REF 1301-P3	AccuDiag™ - Rubella IgG ELISA
Revision Date: 2012-11-15	